

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU071124\
 Data File : VU059897.D
 Acq On : 11 Jul 2024 09:44
 Operator : MD/SY
 Sample : VU0711WBL01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK092

Quant Time: Jul 12 02:04:11 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM070224WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jul 12 02:03:36 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	218043	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	225252	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	94989	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.592	65	71664	51.189	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	= 102.380%	
7) Chloroethane-d5	1.904	69	62594	53.155	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	= 106.320%	
11) 1,1-Dichloroethene-d2	2.557	63	91801	36.910	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	= 73.820%	
21) 2-Butanone-d5	4.615	46	96984	99.682	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	= 99.680%	
24) Chloroform-d	5.052	84	140177	51.705	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	= 103.400%	
26) 1,2-Dichloroethane-d4	5.692	65	87500	54.281	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	= 108.560%	
32) Benzene-d6	5.718	84	285751	49.540	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	= 99.080%	
36) 1,2-Dichloropropane-d6	6.679	67	91702	51.368	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	= 102.740%	
41) Toluene-d8	7.891	98	244757	44.621	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	= 89.240%	
43) trans-1,3-Dichloroprop...	8.174	79	38305	45.854	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	= 91.700%	
47) 2-Hexanone-d5	8.627	63	67569	89.810	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	= 89.810%	
56) 1,1,2,2-Tetrachloroeth...	10.750	84	142404	51.971	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	= 103.940%	
66) 1,2-Dichlorobenzene-d4	12.187	152	92798	53.164	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	= 106.320%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU071124\
Data File : VU059897.D
Acq On : 11 Jul 2024 09:44
Operator : MD/SY
Sample : VU0711WBL01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK092

Quant Time: Jul 12 02:04:11 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM070224WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Jul 12 02:03:36 2024
Response via : Initial Calibration

