

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU071125\
 Data File : VU063484.D
 Acq On : 11 Jul 2025 09:39
 Operator : MD/SY
 Sample : VU0711WBL01
 Misc : 5mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VU0711WBL01

Quant Time: Jul 12 03:44:35 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U070825W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jul 12 03:41:46 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.360	168	111356	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.238	114	199121	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.409	117	182276	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.804	152	86625	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.692	65	75524	53.157	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	106.320%
35) Dibromofluoromethane	5.277	113	59435	49.518	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.040%
50) Toluene-d8	7.888	98	190892	48.790	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.580%
62) 4-Bromofluorobenzene	10.624	95	79131	46.239	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.480%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU071125\
Data File : VU063484.D
Acq On : 11 Jul 2025 09:39
Operator : MD/SY
Sample : VU0711WBL01
Misc : 5mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU0711WBL01

Quant Time: Jul 12 03:44:35 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U070825W.M
Quant Title : SW846 8260
QLast Update : Sat Jul 12 03:41:46 2025
Response via : Initial Calibration

