

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU071218\
 Data File : VU025345.D
 Acq On : 12 Jul 2018 17:32
 Operator : MD/SY
 Sample : PB111044TB
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PB111044TB

Quant Time: Jul 13 02:01:30 2018
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U070918W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 10 09:34:16 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	200201	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	304794	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	278888	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	144559	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	150509	48.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.78%	
35) Dibromofluoromethane	4.89	113	118804	46.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.42%	
50) Toluene-d8	7.57	98	428703	48.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.38%	
62) 4-Bromofluorobenzene	10.31	95	158818	44.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.60%	
Target Compounds						
16) Acetone	2.32	43	12679	12.700	ug/l	Qvalue 100
18) Methyl Acetate	2.62	43	16563	5.201	ug/l	98
20) Methylene Chloride	2.71	84	5814	3.075	ug/l	94
95) Naphthalene	13.75	128	8778	3.200	ug/l #	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU071218\
Data File : VU025345.D
Acq On : 12 Jul 2018 17:32
Operator : MD/SY
Sample : PB111044TB
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PB111044TB

Quant Time: Jul 13 02:01:30 2018
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U070918W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 10 09:34:16 2018
Response via : Initial Calibration

