

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU071218\
 Data File : VU025348.D
 Acq On : 12 Jul 2018 18:45
 Operator : MD/SY
 Sample : PB111044ZHE#03
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PB111044ZHE#03

Quant Time: Jul 13 02:23:23 2018
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U070918W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 10 09:34:16 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	200181	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	308438	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	284165	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	150852	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	152808	49.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.26%	
35) Dibromofluoromethane	4.89	113	120279	46.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.46%	
50) Toluene-d8	7.57	98	435258	48.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.70%	
62) 4-Bromofluorobenzene	10.31	95	162438	45.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.56%	
Target Compounds						
16) Acetone	2.32	43	13148	12.991	ug/l	97
18) Methyl Acetate	2.62	43	18606	5.843	ug/l	98
20) Methylene Chloride	2.71	84	6060	3.176	ug/l	93
95) Naphthalene	13.75	128	5690	2.902	ug/l #	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU071218\
Data File : VU025348.D
Acq On : 12 Jul 2018 18:45
Operator : MD/SY
Sample : PB111044ZHE#03
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PB111044ZHE#03

Quant Time: Jul 13 02:23:23 2018
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U070918W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 10 09:34:16 2018
Response via : Initial Calibration

