

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU071218\
 Data File : VU025364.D
 Acq On : 13 Jul 2018 02:01
 Operator : MD/SY
 Sample : PB111044ZHE#02
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PB111044ZHE#02

Quant Time: Jul 13 05:05:24 2018
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U070918W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 10 09:34:16 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	199388	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	305807	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	279260	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	146803	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	153396	49.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.04%	
35) Dibromofluoromethane	4.89	113	118108	45.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.56%	
50) Toluene-d8	7.57	98	433356	49.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.12%	
62) 4-Bromofluorobenzene	10.31	95	158899	44.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.36%	
Target Compounds						
16) Acetone	2.32	43	12271	12.479	ug/l	95
18) Methyl Acetate	2.62	43	16880	5.322	ug/l	99
20) Methylene Chloride	2.70	84	6162	3.227	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU071218\
Data File : VU025364.D
Acq On : 13 Jul 2018 02:01
Operator : MD/SY
Sample : PB111044ZHE#02
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PB111044ZHE#02

Quant Time: Jul 13 05:05:24 2018
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U070918W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 10 09:34:16 2018
Response via : Initial Calibration

