

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU071218\
 Data File : VU025365.D
 Acq On : 13 Jul 2018 02:25
 Operator : MD/SY
 Sample : PB111044TB
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PB111044TB

Quant Time: Jul 13 05:10:52 2018
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U070918W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 10 09:34:16 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	194299	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	301467	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	278228	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	143642	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	152681	50.58	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.16%	
35) Dibromofluoromethane	4.89	113	116882	45.96	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.92%	
50) Toluene-d8	7.57	98	434863	49.93	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.86%	
62) 4-Bromofluorobenzene	10.31	95	158050	45.08	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.16%	
Target Compounds						
16) Acetone	2.32	43	9839	11.130	ug/l	97
18) Methyl Acetate	2.62	43	16882	5.462	ug/l	100
20) Methylene Chloride	2.71	84	6470	3.424	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU071218\
 Data File : VU025365.D
 Acq On : 13 Jul 2018 02:25
 Operator : MD/SY
 Sample : PB111044TB
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PB111044TB

Quant Time: Jul 13 05:10:52 2018
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U070918W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 10 09:34:16 2018
 Response via : Initial Calibration

