

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU071218\
 Data File : VU025366.D
 Acq On : 13 Jul 2018 02:50
 Operator : MD/SY
 Sample : J3927-04
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 CS-RBR-200031

Quant Time: Jul 13 05:21:11 2018
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U070918W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 10 09:34:16 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	193734	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	298205	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	275720	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	146009	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	150602	50.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.08%	
35) Dibromofluoromethane	4.89	113	114247	45.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.84%	
50) Toluene-d8	7.57	98	423541	49.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.34%	
62) 4-Bromofluorobenzene	10.31	95	158685	45.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.50%	
Target Compounds						
8) Diethyl Ether	2.10	74	1830	1.408	ug/l	96
16) Acetone	2.32	43	15341	14.661	ug/l	94
18) Methyl Acetate	2.62	43	17773	5.767	ug/l	96
20) Methylene Chloride	2.71	84	13206	6.285	ug/l	91
30) Chloroform	4.69	83	4545	1.046	ug/l	94
95) Naphthalene	13.75	128	163903	16.964	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU071218\
Data File : VU025366.D
Acq On : 13 Jul 2018 02:50
Operator : MD/SY
Sample : J3927-04
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
CS-RBR-200031

Quant Time: Jul 13 05:21:11 2018
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U070918W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 10 09:34:16 2018
Response via : Initial Calibration

