

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU071218\
 Data File : VU025367.D
 Acq On : 13 Jul 2018 03:14
 Operator : MD/SY
 Sample : J3928-02
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 SOIL-CUTTINGS

Quant Time: Jul 13 05:25:55 2018
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U070918W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 10 09:34:16 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	191500	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	297926	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	274810	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	143860	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	150619	50.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.24%	
35) Dibromofluoromethane	4.89	113	115771	46.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.12%	
50) Toluene-d8	7.57	98	428716	49.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.62%	
62) 4-Bromofluorobenzene	10.31	95	155917	45.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.00%	
Target Compounds						
8) Diethyl Ether	2.11	74	1609	1.253	ug/l	73
16) Acetone	2.32	43	16919	15.795	ug/l	98
18) Methyl Acetate	2.62	43	17056	5.599	ug/l	99
20) Methylene Chloride	2.71	84	21339	9.835	ug/l	98
30) Chloroform	4.68	83	4455	1.038	ug/l	97
95) Naphthalene	13.75	128	83712	9.956	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU071218\
Data File : VU025367.D
Acq On : 13 Jul 2018 03:14
Operator : MD/SY
Sample : J3928-02
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
SOIL-CUTTINGS

Quant Time: Jul 13 05:25:55 2018
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U070918W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 10 09:34:16 2018
Response via : Initial Calibration

