

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU071218\
 Data File : VU025373.D
 Acq On : 13 Jul 2018 05:39
 Operator : MD/SY
 Sample : J3832-02
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 TR-05-071018

Quant Time: Jul 13 06:41:23 2018
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U070918W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 10 09:34:16 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	187992	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	291824	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	269111	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	145129	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	146635	50.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.42%	
35) Dibromofluoromethane	4.89	113	113267	46.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.02%	
50) Toluene-d8	7.57	98	414246	49.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.28%	
62) 4-Bromofluorobenzene	10.31	95	153729	45.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.58%	
Target Compounds						
16) Acetone	2.32	43	15753	15.231	ug/l	98
18) Methyl Acetate	2.62	43	18796	6.286	ug/l	99
20) Methylene Chloride	2.71	84	100027	44.348	ug/l	99
30) Chloroform	4.68	83	5697	1.352	ug/l	95
95) Naphthalene	13.74	128	157241	16.457	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU071218\
Data File : VU025373.D
Acq On : 13 Jul 2018 05:39
Operator : MD/SY
Sample : J3832-02
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
TR-05-071018

Quant Time: Jul 13 06:41:23 2018
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U070918W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 10 09:34:16 2018
Response via : Initial Calibration

