

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU071218\
 Data File : VU025374.D
 Acq On : 13 Jul 2018 06:03
 Operator : MD/SY
 Sample : J3926-02
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 TRAP-BAG

Quant Time: Jul 13 06:43:39 2018
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U070918W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 10 09:34:16 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	188461	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	292450	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	272727	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	145520	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	146500	50.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.06%	
35) Dibromofluoromethane	4.89	113	111628	45.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.50%	
50) Toluene-d8	7.57	98	416121	49.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.52%	
62) 4-Bromofluorobenzene	10.31	95	154685	45.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.96%	
Target Compounds						
8) Diethyl Ether	2.10	74	1869	1.479	ug/l	83
16) Acetone	2.32	43	13024	13.415	ug/l	98
18) Methyl Acetate	2.62	43	17997	6.003	ug/l	97
20) Methylene Chloride	2.71	84	20714	9.710	ug/l	97
95) Naphthalene	13.74	128	157100	16.407	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU071218\
Data File : VU025374.D
Acq On : 13 Jul 2018 06:03
Operator : MD/SY
Sample : J3926-02
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
TRAP-BAG

Quant Time: Jul 13 06:43:39 2018
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U070918W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 10 09:34:16 2018
Response via : Initial Calibration

