

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU071218\
 Data File : VU025375.D
 Acq On : 13 Jul 2018 06:27
 Operator : MD/SY
 Sample : J3947-02
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PAR-1

Quant Time: Jul 13 06:57:31 2018
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U070918W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 10 09:34:16 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	192179	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	300691	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	277558	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	146794	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	151947	50.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.78%	
35) Dibromofluoromethane	4.89	113	118437	46.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.38%	
50) Toluene-d8	7.57	98	430047	49.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.02%	
62) 4-Bromofluorobenzene	10.31	95	157377	45.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.00%	
Target Compounds						
16) Acetone	2.32	43	17177	15.922	ug/l	97
18) Methyl Acetate	2.62	43	19695	6.443	ug/l	95
20) Methylene Chloride	2.71	84	95668	41.536	ug/l	99
95) Naphthalene	13.74	128	176982	18.041	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU071218\
Data File : VU025375.D
Acq On : 13 Jul 2018 06:27
Operator : MD/SY
Sample : J3947-02
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PAR-1

Quant Time: Jul 13 06:57:31 2018
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U070918W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 10 09:34:16 2018
Response via : Initial Calibration

