


Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU071218\
Data File : VU025331.D
Acq On : 12 Jul 2018 11:04
Operator : MD/SY
Sample : BFB
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U070918W.M
Title : SW846 8260
Last Update : Tue Jul 10 09:34:16 2018



AutoFind: Scans 3023, 3024, 3025; Background Corrected with Scan 3010

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.2	7193	PASS
75	95	30	60	51.2	19189	PASS
95	95	100	100	100.0	37501	PASS
96	95	5	9	6.9	2579	PASS
173	174	0.00	2	1.2	389	PASS
174	95	50	100	85.0	31858	PASS
175	174	5	9	6.9	2196	PASS
176	174	95	101	95.7	30490	PASS
177	176	5	9	6.9	2098	PASS