

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU071320\
 Data File : VU039470.D
 Acq On : 13 Jul 2020 15:11
 Operator : SY/MD
 Sample : L3244-10
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK01

Quant Time: Jul 14 07:28:17 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR070720WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Jul 14 07:08:13 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.27	114	99680	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	93917	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	43296	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	34669	6.00	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	120.00%
7) Chloroethane-d5	1.93	69	30839	5.22	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	104.40%
11) 1,1-Dichloroethene-d2	2.58	63	48680	4.28	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	85.60%
20) 2-Butanone-d5	4.66	46	129483	48.98	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	97.96%
24) Chloroform-d	5.09	84	65399	5.32	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	106.40%
26) 1,2-Dichloroethane-d4	5.72	65	40628	5.26	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	105.20%
32) Benzene-d6	5.75	84	131848	5.83	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	116.60%
36) 1,2-Dichloropropane-d6	6.71	67	42819	5.39	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	107.80%
41) Toluene-d8	7.91	98	119118	6.08	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	121.60%
43) trans-1,3-Dichloropropene-	8.19	79	18064	5.15	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	103.00%
46) 2-Hexanone-d5	8.64	63	96447	46.35	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	92.70%
57) 1,1,2,2-Tetrachloroethane-	10.76	84	38035	4.83	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	96.60%
64) 1,2-Dichlorobenzene-d4	12.19	152	43467	5.62	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	112.40%

Target Compounds					Ovalue
16) Methylene chloride	3.06	84	1689	0.227 ug/L	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU071320\
Data File : VU039470.D
Acq On : 13 Jul 2020 15:11
Operator : SY/MD
Sample : L3244-10
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK01

Quant Time: Jul 14 07:28:17 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR070720WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Jul 14 07:08:13 2020
Response via : Initial Calibration

