

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU071320\
 Data File : VU039468.D
 Acq On : 13 Jul 2020 14:06
 Operator : SY/MD
 Sample : VU0713WBL01
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK05

Quant Time: Jul 14 07:18:03 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR070720WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Jul 14 07:08:13 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	102077	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	94620	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	44222	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	33876	5.73	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	114.60%
7) Chloroethane-d5	1.93	69	31674	5.23	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	104.60%
11) 1,1-Dichloroethene-d2	2.58	63	48815	4.19	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	83.80%
20) 2-Butanone-d5	4.66	46	123726	45.71	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	91.42%
24) Chloroform-d	5.08	84	64232	5.10	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.00%
26) 1,2-Dichloroethane-d4	5.72	65	40351	5.10	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	102.00%
32) Benzene-d6	5.75	84	129051	5.67	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	113.40%
36) 1,2-Dichloropropane-d6	6.71	67	41953	5.24	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	104.80%
41) Toluene-d8	7.91	98	116338	5.89	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	117.80%
43) trans-1,3-Dichloropropene-	8.19	79	17388	4.92	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	98.40%
46) 2-Hexanone-d5	8.64	63	90040	42.95	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	85.90%
57) 1,1,2,2-Tetrachloroethane-	10.76	84	35805	4.51	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	90.20%
64) 1,2-Dichlorobenzene-d4	12.19	152	42147	5.34	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	106.80%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU071320\
Data File : VU039468.D
Acq On : 13 Jul 2020 14:06
Operator : SY/MD
Sample : VU0713WBL01
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK05

Quant Time: Jul 14 07:18:03 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR070720WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Jul 14 07:08:13 2020
Response via : Initial Calibration

