

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU071320\
 Data File : VU039469.D
 Acq On : 13 Jul 2020 14:48
 Operator : SY/MD
 Sample : L3231-10
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK01

Quant Time: Jul 14 07:25:11 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR070720WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Jul 14 07:08:13 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	99175	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	93523	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	43144	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	30453	5.30	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	106.00%
7) Chloroethane-d5	1.92	69	27311	4.64	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	92.80%
11) 1,1-Dichloroethene-d2	2.58	63	45801	4.05	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	81.00%
20) 2-Butanone-d5	4.66	46	119662	45.50	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	91.00%
24) Chloroform-d	5.08	84	61546	5.03	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.60%
26) 1,2-Dichloroethane-d4	5.72	65	38021	4.95	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.00%
32) Benzene-d6	5.74	84	123613	5.49	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	109.80%
36) 1,2-Dichloropropane-d6	6.70	67	39714	5.02	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	100.40%
41) Toluene-d8	7.91	98	110148	5.64	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	112.80%
43) trans-1,3-Dichloropropene-	8.19	79	15928	4.56	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	91.20%
46) 2-Hexanone-d5	8.64	63	86502	41.75	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	83.50%
57) 1,1,2,2-Tetrachloroethane-	10.76	84	35431	4.52	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	90.40%
64) 1,2-Dichlorobenzene-d4	12.19	152	40541	5.26	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	105.20%


Target Compounds					Ovalue
16) Methylene chloride	3.06	84	1952	0.263 ug/L	80


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU071320\
Data File : VU039469.D
Acq On : 13 Jul 2020 14:48
Operator : SY/MD
Sample : L3231-10
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK01

Quant Time: Jul 14 07:25:11 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR070720WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Jul 14 07:08:13 2020
Response via : Initial Calibration





#16
Methylene chloride
Concen: 0.263 ug/L
RT: 3.06 min Scan# 536
Delta R.T. 0.00 min
Lab File: VU039469.D
Acq: 13 Jul 2020 14:48

Instrument :
MSVOA_U
ClientSampleId :
VHBLK01

Tot Ion: 84 Resp: 1952
Ion Ratio Lower Upper
84 100
86 43.7 43.5 80.9
49 109.6 91.1 169.1

