

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU071320\
 Data File : VU039473.D
 Acq On : 13 Jul 2020 16:20
 Operator : SY/MD
 Sample : VIBLK01
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VIBLK01

Quant Time: Jul 14 07:45:56 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR070720WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Jul 14 07:08:13 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.27	114	94367	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	90763	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	42093	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	31359	5.74	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	114.80%
7) Chloroethane-d5	1.93	69	32642	5.83	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	116.60%
11) 1,1-Dichloroethene-d2	2.58	63	46741	4.34	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	86.80%
20) 2-Butanone-d5	4.65	46	118625	47.40	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	94.80%
24) Chloroform-d	5.09	84	57920	4.97	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.40%
26) 1,2-Dichloroethane-d4	5.72	65	37125	5.08	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.60%
32) Benzene-d6	5.74	84	120786	5.53	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	110.60%
36) 1,2-Dichloropropane-d6	6.71	67	40092	5.22	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	104.40%
41) Toluene-d8	7.91	98	108078	5.71	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	114.20%
43) trans-1,3-Dichloropropene-	8.19	79	16008	4.72	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	94.40%
46) 2-Hexanone-d5	8.64	63	87121	43.32	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	86.64%
57) 1,1,2,2-Tetrachloroethane-	10.76	84	34438	4.53	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	90.60%
64) 1,2-Dichlorobenzene-d4	12.19	152	39873	5.31	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	106.20%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU071320\
Data File : VU039473.D
Acq On : 13 Jul 2020 16:20
Operator : SY/MD
Sample : VIBLK01
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VIBLK01

Quant Time: Jul 14 07:45:56 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR070720WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Jul 14 07:08:13 2020
Response via : Initial Calibration

