

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU071323\
 Data File : VU054784.D
 Acq On : 13 Jul 2023 10:08
 Operator : MD/SY
 Sample : VU0713WBL01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK017

Quant Time: Jul 14 07:53:50 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR070723WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Jul 14 07:53:34 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.248	114	217161	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	211505	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.810	152	96250	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	77070	4.703	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	94.000%
7) Chloroethane-d5	1.911	69	68487	4.995	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	99.800%
11) 1,1-Dichloroethene-d2	2.563	65	32207	4.812	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	96.200%
20) 2-Butanone-d5	4.627	46	206899	55.558	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	111.120%
24) Chloroform-d	5.058	84	153146	5.245	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	105.000%
26) 1,2-Dichloroethane-d4	5.698	65	83582	5.476	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	109.600%
32) Benzene-d6	5.724	84	290282	5.170	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	103.400%
36) 1,2-Dichloropropane-d6	6.688	67	93550	5.254	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	105.000%
41) Toluene-d8	7.894	98	256764	4.907	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.200%
43) trans-1,3-Dichloroprop...	8.177	79	36429	5.084	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	101.600%
46) 2-Hexanone-d5	8.630	63	147379	53.686	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	107.380%
56) 1,1,2,2-Tetrachloroeth...	10.753	84	73235	5.104	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	102.000%
66) 1,2-Dichlorobenzene-d4	12.190	152	92025	5.488	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	109.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU071323\
Data File : VU054784.D
Acq On : 13 Jul 2023 10:08
Operator : MD/SY
Sample : VU0713WBL01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK017

Quant Time: Jul 14 07:53:50 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR070723WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Jul 14 07:53:34 2023
Response via : Initial Calibration

