

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU071422\
 Data File : VU049713.D
 Acq On : 14 Jul 2022 12:43
 Operator : SY/MD
 Sample : N3676-10ME
 Misc : 6.44g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 F5B09ME

Quant Time: Jul 15 04:02:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM071122WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jul 15 04:01:16 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.243	114	258760	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	252692	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.815	152	120591	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.594	65	69164	31.679	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	63.360%
7) Chloroethane-d5	1.874	69	41497	26.295	ug/L	-0.03
Spiked Amount	50.000	Range	70 - 130	Recovery	=	52.580%#
11) 1,1-Dichloroethene-d2	2.520	63	149375	38.733	ug/L	-0.04
Spiked Amount	50.000	Range	60 - 125	Recovery	=	77.460%
21) 2-Butanone-d5	4.636	46	281379	150.971	ug/L	0.02
Spiked Amount	100.000	Range	40 - 130	Recovery	=	150.970%#
24) Chloroform-d	5.060	84	241645	62.490	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	124.980%
26) 1,2-Dichloroethane-d4	5.697	65	174829	66.802	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	133.600%#
32) Benzene-d6	5.719	84	466440	60.760	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	121.520%
36) 1,2-Dichloropropane-d6	6.687	67	166073	64.317	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	128.640%#
41) Toluene-d8	7.896	98	400281	59.875	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	119.760%
43) trans-1,3-Dichloroprop...	8.179	79	67877	62.392	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	124.780%
47) 2-Hexanone-d5	8.655	63	187269	161.230	ug/L	0.02
Spiked Amount	100.000	Range	45 - 130	Recovery	=	161.230%#
56) 1,1,2,2-Tetrachloroeth...	10.764	84	260960	72.229	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	144.460%#
66) 1,2-Dichlorobenzene-d4	12.195	152	155798	68.522	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	137.040%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU071422\
Data File : VU049713.D
Acq On : 14 Jul 2022 12:43
Operator : SY/MD
Sample : N3676-10ME
Misc : 6.44g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
F5B09ME

Quant Time: Jul 15 04:02:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM071122WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Jul 15 04:01:16 2022
Response via : Initial Calibration

