

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU071422\
 Data File : VU049714.D
 Acq On : 14 Jul 2022 13:07
 Operator : SY/MD
 Sample : N3678-01ME
 Misc : 5.99g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 F5B20ME

Quant Time: Jul 15 04:02:59 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM071122WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jul 15 04:01:16 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.243	114	218792	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	208963	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.815	152	98431	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.594	65	40862	22.135	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	44.260%#
7) Chloroethane-d5	1.880	69	28874	21.638	ug/L	-0.02
Spiked Amount	50.000	Range	70 - 130	Recovery	=	43.280%#
11) 1,1-Dichloroethene-d2	2.526	63	93298	28.612	ug/L	-0.04
Spiked Amount	50.000	Range	60 - 125	Recovery	=	57.220%#
21) 2-Butanone-d5	4.639	46	171860	109.054	ug/L	0.02
Spiked Amount	100.000	Range	40 - 130	Recovery	=	109.050%
24) Chloroform-d	5.060	84	147524	45.119	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.240%
26) 1,2-Dichloroethane-d4	5.700	65	104132	47.057	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.120%
32) Benzene-d6	5.719	84	272858	42.982	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	85.960%
36) 1,2-Dichloropropane-d6	6.687	67	100093	46.876	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	93.760%
41) Toluene-d8	7.896	98	230655	41.722	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	83.440%
43) trans-1,3-Dichloroprop...	8.182	79	38335	42.611	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	85.220%
47) 2-Hexanone-d5	8.655	63	109975	114.497	ug/L	0.02
Spiked Amount	100.000	Range	45 - 130	Recovery	=	114.500%
56) 1,1,2,2-Tetrachloroeth...	10.764	84	157914	52.854	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	105.700%
66) 1,2-Dichlorobenzene-d4	12.195	152	95470	51.442	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	102.880%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU071422\
Data File : VU049714.D
Acq On : 14 Jul 2022 13:07
Operator : SY/MD
Sample : N3678-01ME
Misc : 5.99g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
F5B20ME

Quant Time: Jul 15 04:02:59 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM071122WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Jul 15 04:01:16 2022
Response via : Initial Calibration

