

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU071422\
 Data File : VU049716.D
 Acq On : 14 Jul 2022 13:54
 Operator : SY/MD
 Sample : N3676-10ME
 Misc : 6.44g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 F5B09ME

Quant Time: Jul 15 04:03:28 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM071122WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jul 15 04:01:16 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.243	114	215372	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	210564	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.815	152	97275	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.594	65	39177	21.559	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	43.120%#
7) Chloroethane-d5	1.871	69	23639	17.996	ug/L	-0.03
Spiked Amount	50.000	Range	70 - 130	Recovery	=	36.000%#
11) 1,1-Dichloroethene-d2	2.523	63	85148	26.527	ug/L	-0.04
Spiked Amount	50.000	Range	60 - 125	Recovery	=	53.060%#
21) 2-Butanone-d5	4.636	46	165423	106.637	ug/L	0.02
Spiked Amount	100.000	Range	40 - 130	Recovery	=	106.640%
24) Chloroform-d	5.057	84	138636	43.074	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	86.140%
26) 1,2-Dichloroethane-d4	5.697	65	99191	45.536	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.080%
32) Benzene-d6	5.719	84	253366	39.608	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	79.220%
36) 1,2-Dichloropropane-d6	6.687	67	92654	43.062	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	86.120%
41) Toluene-d8	7.896	98	216370	38.841	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	77.680%#
43) trans-1,3-Dichloroprop...	8.179	79	35865	39.563	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	79.120%
47) 2-Hexanone-d5	8.648	63	107074	110.629	ug/L	0.02
Spiked Amount	100.000	Range	45 - 130	Recovery	=	110.630%
56) 1,1,2,2-Tetrachloroeth...	10.761	84	150679	50.049	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	100.100%
66) 1,2-Dichlorobenzene-d4	12.195	152	89208	48.639	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.280%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU071422\
Data File : VU049716.D
Acq On : 14 Jul 2022 13:54
Operator : SY/MD
Sample : N3676-10ME
Misc : 6.44g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
F5B09ME

Quant Time: Jul 15 04:03:28 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM071122WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Jul 15 04:01:16 2022
Response via : Initial Calibration

