

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU071425\
 Data File : VU063502.D
 Acq On : 14 Jul 2025 11:02
 Operator : MD/SY
 Sample : VU0714WBL02
 Misc : 5mL/MSVOA_U/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VU0714WBL02

Quant Time: Jul 15 01:31:37 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U070825W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 15 01:29:09 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.361	168	102635	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.238	114	181957	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.409	117	167595	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.807	152	85994	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.692	65	70611	53.922	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	107.840%
35) Dibromofluoromethane	5.280	113	54882	50.038	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.080%
50) Toluene-d8	7.888	98	168939	47.253	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.500%
62) 4-Bromofluorobenzene	10.627	95	76840	49.136	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.280%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU071425\
Data File : VU063502.D
Acq On : 14 Jul 2025 11:02
Operator : MD/SY
Sample : VU0714WBL02
Misc : 5mL/MSVOA_U/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU0714WBL02

Quant Time: Jul 15 01:31:37 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U070825W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 15 01:29:09 2025
Response via : Initial Calibration

