

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU071519\
 Data File : VU033248.D
 Acq On : 15 Jul 2019 11:01
 Operator : JC/SP
 Sample : K3760-20
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK02

Quant Time: Jul 16 02:08:34 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM071219WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Jul 16 01:46:18 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.86	114	302376	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.07	117	302363	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.47	152	131825	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.39	65	102291	44.69	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	89.38%
7) Chloroethane-d5	1.67	69	92000	47.05	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	94.10%
11) 1,1-Dichloroethene-d2	2.26	63	134432	32.38	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	64.76%
21) 2-Butanone-d5	4.15	46	150795	103.97	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	103.97%
24) Chloroform-d	4.62	84	180952	47.17	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.34%
26) 1,2-Dichloroethane-d4	5.29	65	121773	49.50	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.00%
32) Benzene-d6	5.32	84	365260	49.24	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.48%
36) 1,2-Dichloropropane-d6	6.31	67	119550	50.31	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	100.62%
41) Toluene-d8	7.55	98	328887	47.02	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	94.04%
43) trans-1,3-Dichloropropene-	7.83	79	56364	48.38	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	96.76%
47) 2-Hexanone-d5	8.29	63	107628	100.48	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	100.48%
57) 1,1,2,2-Tetrachloroethane-	10.41	84	177760	49.43	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	98.86%
64) 1,2-Dichlorobenzene-d4	11.85	152	124604	53.00	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	106.00%

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU071519\
 Data File : VU033248.D
 Acq On : 15 Jul 2019 11:01
 Operator : JC/SP
 Sample : K3760-20
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampled :
 VHBLK02

Quant Time: Jul 16 02:08:34 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM071219WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Jul 16 01:46:18 2019
 Response via : Initial Calibration

