

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\U071520\
 Data File : VU039505.D
 Acq On : 15 Jul 2020 17:29
 Operator : SY/MD
 Sample : L3313-04
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 F4C03

Quant Time: Jul 16 06:04:27 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR070720WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Jul 16 04:09:55 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.27	114	91766	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	86435	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	41990	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	20453	3.85	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	77.00%
7) Chloroethane-d5	1.93	69	25468	4.68	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	93.60%
11) 1,1-Dichloroethene-d2	2.58	63	34513	3.30	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	66.00%
20) 2-Butanone-d5	4.65	46	98427	40.45	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	80.90%
24) Chloroform-d	5.09	84	48063	4.24	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.80%
26) 1,2-Dichloroethane-d4	5.72	65	30766	4.33	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	86.60%
32) Benzene-d6	5.74	84	94136	4.52	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.40%
36) 1,2-Dichloropropane-d6	6.71	67	32789	4.48	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	89.60%
41) Toluene-d8	7.91	98	81673	4.53	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.60%
43) trans-1,3-Dichloropropene-	8.19	79	12224	3.79	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	75.80%
46) 2-Hexanone-d5	8.64	63	69273	36.17	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	72.34%
57) 1,1,2,2-Tetrachloroethane-	10.76	84	28646	3.95	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	79.00%
64) 1,2-Dichlorobenzene-d4	12.19	152	33078	4.41	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	88.20%

Target Compounds					Ovalue
25) Chloroform	5.12	83	2800	0.235 ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU071520\
Data File : VU039505.D
Acq On : 15 Jul 2020 17:29
Operator : SY/MD
Sample : L3313-04
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
F4C03

Quant Time: Jul 16 06:04:27 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR070720WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Jul 16 04:09:55 2020
Response via : Initial Calibration

