

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\U071520\
 Data File : VU039508.D
 Acq On : 15 Jul 2020 18:38
 Operator : SY/MD
 Sample : L3313-07
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 F4C07

Quant Time: Jul 16 06:18:12 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR070720WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Jul 16 04:09:55 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	93488	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	89265	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	42766	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	21370	3.95	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	79.00%
7) Chloroethane-d5	1.92	69	25565	4.61	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	92.20%
11) 1,1-Dichloroethene-d2	2.58	63	33313	3.13	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	62.60%
20) 2-Butanone-d5	4.65	46	98060	39.55	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	79.10%
24) Chloroform-d	5.08	84	49816	4.32	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.40%
26) 1,2-Dichloroethane-d4	5.72	65	30401	4.20	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	84.00%
32) Benzene-d6	5.75	84	95446	4.44	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.80%
36) 1,2-Dichloropropane-d6	6.71	67	32367	4.29	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	85.80%
41) Toluene-d8	7.91	98	82339	4.42	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.40%
43) trans-1,3-Dichloropropene-	8.19	79	12264	3.68	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	73.60%
46) 2-Hexanone-d5	8.64	63	70305	35.55	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	71.10%
57) 1,1,2,2-Tetrachloroethane-	10.76	84	28614	3.82	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	76.40%
64) 1,2-Dichlorobenzene-d4	12.19	152	33217	4.35	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	87.00%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU071520\
Data File : VU039508.D
Acq On : 15 Jul 2020 18:38
Operator : SY/MD
Sample : L3313-07
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
F4C07

Quant Time: Jul 16 06:18:12 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR070720WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Jul 16 04:09:55 2020
Response via : Initial Calibration

