

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU071619\
 Data File : VU033265.D
 Acq On : 16 Jul 2019 16:12
 Operator : JC/SP
 Sample : VIBLK41
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VIBLK41

Quant Time: Jul 17 05:24:58 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR071619WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Jul 17 04:42:24 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.86	114	128359	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.07	117	132642	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.47	152	53216	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.39	65	54519	4.22	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	84.40%
7) Chloroethane-d5	1.67	69	49883	4.63	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	92.60%
11) 1,1-Dichloroethene-d2	2.26	63	84996	3.40	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	68.00%
20) 2-Butanone-d5	4.18	46	117731	48.33	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	96.66%
24) Chloroform-d	4.62	84	88896	4.66	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.20%
26) 1,2-Dichloroethane-d4	5.29	65	48853	4.74	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.80%
32) Benzene-d6	5.32	84	167608	4.54	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.80%
36) 1,2-Dichloropropane-d6	6.31	67	56051	4.77	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	95.40%
41) Toluene-d8	7.55	98	147042	4.20	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	84.00%
43) trans-1,3-Dichloropropene-	7.84	79	20413	4.34	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	86.80%
46) 2-Hexanone-d5	8.30	63	85650	44.27	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	88.54%
57) 1,1,2,2-Tetrachloroethane-	10.42	84	44140	4.52	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	90.40%
64) 1,2-Dichlorobenzene-d4	11.85	152	53882	5.20	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	104.00%

Target Compounds					Ovalue
22) cis-1,2-Dichloroethene	4.20	96	2230	0.227 ug/L	93
34) Trichloroethene	6.16	95	2593	0.261 ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU071619\
Data File : VU033265.D
Acq On : 16 Jul 2019 16:12
Operator : JC/SP
Sample : VIBLK41
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VIBLK41

Quant Time: Jul 17 05:24:58 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR071619WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Jul 17 04:42:24 2019
Response via : Initial Calibration

