

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\U071620\
 Data File : VU039526.D
 Acq On : 16 Jul 2020 16:05
 Operator : SY/MD
 Sample : L3292-07 20X
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BFT83

Quant Time: Jul 17 02:39:13 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR070720WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Jul 17 01:06:47 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.27	114	95757	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	93447	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	43246	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	20051	3.61	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	72.20%
7) Chloroethane-d5	1.93	69	25172	4.43	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	88.60%
11) 1,1-Dichloroethene-d2	2.58	63	34879	3.19	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	63.80%
20) 2-Butanone-d5	4.66	46	106884	42.09	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	84.18%
24) Chloroform-d	5.09	84	50282	4.25	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	85.00%
26) 1,2-Dichloroethane-d4	5.72	65	32212	4.34	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	86.80%
32) Benzene-d6	5.75	84	94532	4.20	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.00%
36) 1,2-Dichloropropane-d6	6.71	67	34318	4.34	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	86.80%
41) Toluene-d8	7.91	98	84440	4.33	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	86.60%
43) trans-1,3-Dichloropropene-	8.19	79	12713	3.64	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	72.80%
46) 2-Hexanone-d5	8.64	63	76796	37.09	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	74.18%
57) 1,1,2,2-Tetrachloroethane-	10.76	84	30198	3.85	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	77.00%
64) 1,2-Dichlorobenzene-d4	12.19	152	32862	4.26	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	85.20%

Target Compounds

					Ovalue
10) 1,1,2-Trichloro-1,2,2-trif	2.59	101	1662	0.299 ug/L	89
12) 1,1-Dichloroethene	2.59	96	2490	0.437 ug/L	# 1
16) Methylene chloride	3.06	84	3898	0.544 ug/L	91
22) cis-1,2-Dichloroethene	4.69	96	2780	0.393 ug/L	82
25) Chloroform	5.10	83	3619	0.291 ug/L	99
34) Trichloroethene	6.56	95	36169	5.008 ug/L	98
47) Tetrachloroethene	8.56	164	118479	23.582 ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU071620\
Data File : VU039526.D
Acq On : 16 Jul 2020 16:05
Operator : SY/MD
Sample : L3292-07 20X
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BFT83

Quant Time: Jul 17 02:39:13 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR070720WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jul 17 01:06:47 2020
Response via : Initial Calibration

