

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU071624\
 Data File : VU059981.D
 Acq On : 16 Jul 2024 17:58
 Operator : MD/SY
 Sample : P3217-06
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK001

Quant Time: Jul 17 02:33:03 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR061724WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Jul 17 02:25:31 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	193915	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	198104	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.810	152	94497	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	54177	4.018	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	80.400%
7) Chloroethane-d5	1.907	69	48829	3.764	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	75.200%
11) 1,1-Dichloroethene-d2	2.560	65	22955	4.103	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	82.000%
20) 2-Butanone-d5	4.627	46	147529	53.116	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	106.240%
24) Chloroform-d	5.058	84	111297	4.021	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	80.400%
26) 1,2-Dichloroethane-d4	5.698	65	57231	4.293	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	85.800%
32) Benzene-d6	5.721	84	219069	4.018	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	80.400%
36) 1,2-Dichloropropane-d6	6.685	67	70052	4.193	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	83.800%
41) Toluene-d8	7.894	98	190437	3.757	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	75.200%
43) trans-1,3-Dichloroprop...	8.177	79	24272	4.640	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	92.800%
46) 2-Hexanone-d5	8.631	63	116778	53.724	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	107.440%
56) 1,1,2,2-Tetrachloroeth...	10.753	84	61627	4.465	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	89.200%
66) 1,2-Dichlorobenzene-d4	12.190	152	76382	4.541	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	90.800%
Target Compounds						
					Qvalue	
16) Methylene chloride	3.039	84	4911	0.301	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU071624\
Data File : VU059981.D
Acq On : 16 Jul 2024 17:58
Operator : MD/SY
Sample : P3217-06
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK001

Quant Time: Jul 17 02:33:03 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR061724WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Jul 17 02:25:31 2024
Response via : Initial Calibration

