

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU071718\
 Data File : VU025459.D
 Acq On : 17 Jul 2018 14:05
 Operator : MD/SY
 Sample : VU0717WBL01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK43

Quant Time: Jul 18 05:56:46 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM071718WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Jul 18 05:54:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.89	114	358820	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	320393	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.49	152	153796	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	79217	46.81	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	93.62%
7) Chloroethane-d5	1.68	69	68814	47.74	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	95.48%
11) 1,1-Dichloroethene-d2	2.27	63	132826	35.75	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	71.50%
21) 2-Butanone-d5	4.18	46	175354	97.39	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	97.39%
24) Chloroform-d	4.65	84	202196	46.06	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.12%
26) 1,2-Dichloroethane-d4	5.31	65	142381	47.66	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.32%
32) Benzene-d6	5.35	84	390675	50.46	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.92%
36) 1,2-Dichloropropane-d6	6.33	67	129633	50.56	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	101.12%
41) Toluene-d8	7.57	98	366719	48.93	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.86%
43) trans-1,3-Dichloropropene-	7.85	79	65111	48.78	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	97.56%
47) 2-Hexanone-d5	8.31	63	129964	97.77	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	97.77%
57) 1,1,2,2-Tetrachloroethane-	10.44	84	181483	47.67	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	95.34%
64) 1,2-Dichlorobenzene-d4	11.86	152	152541	49.49	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.98%


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
37) 1,2-Dichloropropane	6.34	63	14658	5.606	ug/L	# 88


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_U\DATA\VU071718\
Data File : VU025459.D
Acq On : 17 Jul 2018 14:05
Operator : MD/SY
Sample : VU0717WBL01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK43

Quant Time: Jul 18 05:56:46 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM071718WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Jul 18 05:54:25 2018
Response via : Initial Calibration





#37
 1,2-Dichloropropane
 Concen: 5.606 ug/L
 RT: 6.34 min Scan# 1787
 Delta R.T. -0.10 min
 Lab File: VU025459.D
 Acq: 17 Jul 2018 14:05

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK43

Tgt Ion: 63 Resp: 14658
 Ion Ratio Lower Upper
 63 100
 112 0.0 3.2 4.8#

