

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\U071725\
 Data File : VU063530.D
 Acq On : 17 Jul 2025 18:21
 Operator : MD/SY
 Sample : VSTDCCC010
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_U
 LabSampleId :
 VSTDCCC010

Quant Time: Jul 18 04:17:43 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U071625DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Thu Jul 17 03:49:40 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 200%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 i	Fluorobenzene	1.000	1.000	0.0	104	0.00
2 T	Dichlorodifluoromethane	0.295	0.234	20.7	82	0.00
3 t	Chloromethane	0.274	0.231	15.7	89	0.00
4 Rt	Vinyl Chloride	0.349	0.307	12.0	90	0.00
5 T	Bromomethane	0.263	0.232	11.8	99	0.00
6 T	Chloroethane	0.211	0.187	11.4	91	0.00
7 T	Trichlorofluoromethane	0.501	0.454	9.4	93	0.00
8	1,1,2-Trichloro-1,2,2-trifl	0.262	0.238	9.2	93	0.00
9 Rt	1,1-Dichloroethene	0.260	0.237	8.8	93	0.00
10 t	Iodomethane	0.314	0.257	18.2	72	0.00
11 t	Allyl Chloride	0.375	0.331	11.7	90	0.00
12 t	Acrylonitrile	0.065	0.055	15.4	90	0.00
13 T	Acetone	0.052	0.043	17.3	89	0.00
14 T	Carbon Disulfide	0.839	0.661	21.2	80	0.00
15 RT	Methylene Chloride	0.306	0.271	11.4	96	0.00
16 RT	trans-1,2-Dichloroethene	0.295	0.264	10.5	92	0.00
17 t	1,1-Dichloroethane	0.546	0.500	8.4	95	0.00
18 T	2-Butanone	0.081	0.071	12.3	92	0.00
19	Cyclohexane	0.463	0.417	9.9	94	0.00
20	Methylcyclohexane	0.450	0.340	24.4	72	0.00
21 T	2,2-Dichloropropane	0.465	0.342	26.5	78	0.00
22 RT	cis-1,2-Dichloroethene	0.320	0.287	10.3	93	0.00
23 t	Diethyl Ether	0.198	0.189	4.5	95	0.00
24 t	tert-Butyl Alcohol	0.024	0.023	4.2	99	0.00
25 t	Methyl tert-Butyl Ether	0.678	0.632	6.8	95	0.00
26 t	Bromochloromethane	0.134	0.124	7.5	94	0.00
27 t	Chloroform	0.558	0.510	8.6	97	0.00
28 RT	1,1,1-Trichloroethane	0.473	0.432	8.7	93	0.00
29 T	1,1-Dichloropropene	0.438	0.396	9.6	95	0.00
30 RT	Carbon Tetrachloride	0.381	0.345	9.4	91	0.00
31 t	Isopropyl Ether	0.825	0.767	7.0	96	0.00
32	Ethyl-t-butyl ether	0.763	0.711	6.8	96	0.00
33	Tert-Amyl methyl ether	0.663	0.605	8.7	94	0.00
34 t	Propionitrile	0.025	0.023	8.0	102	0.00
35 RT	Benzene	1.178	1.103	6.4	95	0.00
36 RT	1,2-Dichloroethane	0.366	0.348	4.9	96	0.00
37 RT	Trichloroethene	0.301	0.239	20.6	77	0.00
38 Rt	1,2-Dichloropropane	0.276	0.232	15.9	76	0.00
39 t	Methacrylonitrile	0.109	0.093	14.7	95	0.00
40 t	Methyl acrylate	0.159	0.151	5.0	93	0.00
41 t	Tetrahydrofuran	0.057	0.051	10.5	100	0.00
42 t	1-Chlorobutane	0.583	0.536	8.1	94	0.00
43 T	Dibromomethane	0.143	0.122	14.7	82	0.00
44 T	Bromodichloromethane	0.322	0.265	17.7	84	0.00
45 T	4-Methyl-2-Pentanone	0.147	0.134	8.8	92	0.00
46 t	t-1,4-Dichloro-2-butene	0.055	0.046	16.4	76	0.00
47 t	Methyl methacrylate	0.120	0.105	12.5	86	0.00
48 t	Ethyl methacrylate	0.229	0.210	8.3	86	0.00
49 Rt	Toluene	0.643	0.570	11.4	88	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU071725\
 Data File : VU063530.D
 Acq On : 17 Jul 2025 18:21
 Operator : MD/SY
 Sample : VSTDCCC010
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_U
 LabSampleId :
 VSTDCCC010

Quant Time: Jul 18 04:17:43 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U071625DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Thu Jul 17 03:49:40 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 200%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
50 T	t-1,3-Dichloropropene	0.253	0.219	13.4	80	0.00
51 T	cis-1,3-Dichloropropene	0.339	0.287	15.3	82	0.00
52 RT	1,1,2-Trichloroethane	0.192	0.177	7.8	91	0.00
53 t	1,3-Dichloropropane	0.332	0.298	10.2	90	0.00
54 t	2-Hexanone	0.098	0.090	8.2	91	0.00
55 t	Dibromochloromethane	0.203	0.183	9.9	82	0.00
56 T	1,2-Dibromoethane	0.172	0.150	12.8	87	0.00
57 S	4-Bromofluorobenzene	0.379	0.323	14.8	91	0.00
58 RT	Tetrachloroethene	0.275	0.232	15.6	85	0.00
59 Rt	Chlorobenzene	0.739	0.652	11.8	88	0.00
60 T	1,1,1,2-Tetrachloroethane	0.238	0.205	13.9	84	0.00
61 t	Pentachloroethane	0.168	0.148	11.9	83	0.00
62 t	Hexachloroethane	0.166	0.145	12.7	78	0.00
63 Rt	Ethyl Benzene	1.271	1.115	12.3	88	0.00
64 RT	m/p-Xylenes	0.494	0.440	10.9	88	0.00
65 RT	o-Xylene	0.475	0.430	9.5	88	0.00
66 RT	Styrene	0.757	0.704	7.0	89	0.00
67 t	Bromoform	0.104	0.085	18.3	74	0.00
68 S	1,2-Dichlorobenzene-d4	0.350	0.340	2.9	98	0.00
69 T	Isopropylbenzene	1.148	1.041	9.3	88	0.00
70 T	1,1,2,2-Tetrachloroethane	0.234	0.217	7.3	93	0.00
71 T	1,2,3-Trichloropropane	0.190	0.162	14.7	81	0.00
72 t	Bromobenzene	0.296	0.260	12.2	87	0.00
73 t	n-propylbenzene	0.346	0.315	9.0	87	0.00
74 t	2-Chlorotoluene	0.305	0.275	9.8	89	0.00
75 t	1,3,5-Trimethylbenzene	1.097	1.011	7.8	89	0.00
76 t	4-Chlorotoluene	0.310	0.282	9.0	89	0.00
77 t	tert-Butylbenzene	1.048	0.947	9.6	88	0.00
78 t	1,2,4-Trimethylbenzene	1.088	1.004	7.7	89	0.00
79 t	sec-Butylbenzene	1.413	1.299	8.1	89	0.00
80	Nitrobenzene	0.004	0.003	25.0	62	0.00
81 t	p-Isopropyltoluene	1.180	1.104	6.4	90	0.00
82 t	1,3-Dichlorobenzene	0.593	0.524	11.6	87	0.00
83 Rt	1,4-Dichlorobenzene	0.593	0.565	4.7	92	0.00
84 t	n-Butylbenzene	1.094	1.011	7.6	87	0.00
85 Rt	1,2-Dichlorobenzene	0.557	0.509	8.6	89	0.00
86 t	1,2-Dibromo-3-Chloropropane	0.033	0.032	3.0	83	0.00
87 Rt	1,2,4-Trichlorobenzene	0.346	0.332	4.0	90	0.00
88 t	Hexachlorobutadiene	0.197	0.167	15.2	82	0.00
89 t	Naphthalene	0.621	0.534	14.0	82	0.00
90 t	1,2,3-Trichlorobenzene	0.320	0.281	12.2	83	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0