

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU071818\
 Data File : VU025509.D
 Acq On : 18 Jul 2018 20:17
 Operator : MD/SY
 Sample : J4039-03MSD
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 DB1R2MSD

Quant Time: Jul 19 02:03:21 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM071718WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Jul 19 01:07:14 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.89	114	320211	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	291869	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.49	152	136664	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	57506	38.08	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	76.16%
7) Chloroethane-d5	1.68	69	55239	42.94	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	85.88%
11) 1,1-Dichloroethene-d2	2.28	63	151423	45.67	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	91.34%
21) 2-Butanone-d5	4.18	46	141616	88.13	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	88.13%
24) Chloroform-d	4.65	84	169667	43.31	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	86.62%
26) 1,2-Dichloroethane-d4	5.31	65	118591	44.48	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.96%
32) Benzene-d6	5.35	84	317592	45.03	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.06%
36) 1,2-Dichloropropane-d6	6.34	67	108965	46.65	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	93.30%
41) Toluene-d8	7.57	98	296761	43.47	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	86.94%
43) trans-1,3-Dichloropropene-	7.85	79	51459	42.32	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	84.64%
47) 2-Hexanone-d5	8.31	63	102438	84.59	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	84.59%
57) 1,1,2,2-Tetrachloroethane-	10.44	84	155458	44.83	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	89.66%
64) 1,2-Dichlorobenzene-d4	11.86	152	127789	46.66	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	93.32%

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU071818\
Data File : VU025509.D
Acq On : 18 Jul 2018 20:17
Operator : MD/SY
Sample : J4039-03MSD
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
DB1R2MSD

Quant Time: Jul 19 02:03:21 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM071718WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Jul 19 01:07:14 2018
Response via : Initial Calibration

