

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU071819\
 Data File : VU033318.D
 Acq On : 18 Jul 2019 14:46
 Operator : JC/SP
 Sample : K3831-01
 Misc : 16.92g/5mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GAMR0

Quant Time: Jul 22 16:13:15 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM071219WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jul 19 01:43:14 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.86	114	384183	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.07	117	383589	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.47	152	223374	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.39	65	126730	43.58	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	87.16%
7) Chloroethane-d5	1.63	69	98754	39.75	ug/L	-0.04
Spiked Amount	50.000	Range	70 - 130	Recovery	=	79.50%
11) 1,1-Dichloroethene-d2	2.22	63	212767	40.33	ug/L	-0.03
Spiked Amount	50.000	Range	60 - 125	Recovery	=	80.66%
21) 2-Butanone-d5	4.16	46	195036	105.84	ug/L	0.01
Spiked Amount	100.000	Range	40 - 130	Recovery	=	105.84%
24) Chloroform-d	4.62	84	222291	45.61	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.22%
26) 1,2-Dichloroethane-d4	5.28	65	143511	45.91	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.82%
32) Benzene-d6	5.31	84	477518	50.74	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.48%
36) 1,2-Dichloropropane-d6	6.31	67	149954	49.75	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	99.50%
41) Toluene-d8	7.55	98	485782	54.74	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	109.48%
43) trans-1,3-Dichloropropene-	7.83	79	80881	54.72	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	109.44%
47) 2-Hexanone-d5	8.31	63	164135	120.78	ug/L	0.01
Spiked Amount	100.000	Range	45 - 130	Recovery	=	120.78%
57) 1,1,2,2-Tetrachloroethane-	10.42	84	244600	53.61	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	107.22%
64) 1,2-Dichlorobenzene-d4	11.85	152	215219	54.03	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	108.06%

Target Compounds

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
35) Methylcyclohexane	6.39	83	625991	139.682	ug/L	93
52) Ethylbenzene	9.23	91	450877	34.938	ug/L	100
53) m,p-Xylene	9.36	106	331381	67.793	ug/L	98
56) Isopropylbenzene	10.15	105	243188	19.471	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU071819\
Data File : VU033318.D
Acq On : 18 Jul 2019 14:46
Operator : JC/SP
Sample : K3831-01
Misc : 16.92g/5mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GAMR0

Quant Time: Jul 22 16:13:15 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM071219WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Jul 19 01:43:14 2019
Response via : Initial Calibration

