

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU071822\
 Data File : VU049758.D
 Acq On : 18 Jul 2022 14:04
 Operator : SY/MD
 Sample : N3748-06
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BH0E8

Quant Time: Jul 18 14:26:11 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM071122WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jul 15 23:03:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.247	114	181659	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	173681	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	70666	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	48663	31.749	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	63.500%	
7) Chloroethane-d5	1.903	69	43144	38.941	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	77.880%	
11) 1,1-Dichloroethene-d2	2.562	63	67337	24.871	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	49.740%#	
21) 2-Butanone-d5	4.620	46	165334	126.359	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	126.360%	
24) Chloroform-d	5.063	84	131280	48.358	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	96.720%	
26) 1,2-Dichloroethane-d4	5.700	65	94383	51.370	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	102.740%	
32) Benzene-d6	5.726	84	232155	43.999	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	88.000%	
36) 1,2-Dichloropropane-d6	6.690	67	91534	51.576	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	103.160%	
41) Toluene-d8	7.896	98	183299	39.892	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	79.780%#	
43) trans-1,3-Dichloroprop...	8.179	79	34737	46.456	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	92.920%	
47) 2-Hexanone-d5	8.632	63	86365	108.182	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	108.180%	
56) 1,1,2,2-Tetrachloroeth...	10.754	84	140866	56.726	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	113.460%	
66) 1,2-Dichlorobenzene-d4	12.192	152	70578	52.972	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	105.940%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU071822\
Data File : VU049758.D
Acq On : 18 Jul 2022 14:04
Operator : SY/MD
Sample : N3748-06
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BH0E8

Quant Time: Jul 18 14:26:11 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM071122WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Jul 15 23:03:46 2022
Response via : Initial Calibration

