

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU071822\
 Data File : VU049774.D
 Acq On : 18 Jul 2022 20:46
 Operator : SY/MD
 Sample : N3700-16ME
 Misc : 6.34g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 F5B57ME

Quant Time: Jul 19 02:14:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM071122WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Jul 19 02:09:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.244	114	199371	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.421	117	192446	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.816	152	90251	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.594	65	25222	14.994	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	29.980%#
7) Chloroethane-d5	1.877	69	19933	16.393	ug/L	-0.02
Spiked Amount	50.000	Range	70 - 130	Recovery	=	32.780%#
11) 1,1-Dichloroethene-d2	2.524	63	55244	18.592	ug/L	-0.04
Spiked Amount	50.000	Range	60 - 125	Recovery	=	37.180%#
21) 2-Butanone-d5	4.636	46	153783	107.089	ug/L	0.02
Spiked Amount	100.000	Range	40 - 130	Recovery	=	107.090%
24) Chloroform-d	5.061	84	113284	38.022	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	76.040%
26) 1,2-Dichloroethane-d4	5.700	65	83471	41.395	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	82.780%
32) Benzene-d6	5.720	84	193717	33.134	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	66.260%#
36) 1,2-Dichloropropane-d6	6.687	67	77071	39.192	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	78.380%
41) Toluene-d8	7.896	98	154261	30.298	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	60.600%#
43) trans-1,3-Dichloroprop...	8.183	79	30003	36.212	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	72.420%
47) 2-Hexanone-d5	8.658	63	95593	108.066	ug/L	0.03
Spiked Amount	100.000	Range	45 - 130	Recovery	=	108.070%
56) 1,1,2,2-Tetrachloroeth...	10.764	84	136634	49.657	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	99.320%
66) 1,2-Dichlorobenzene-d4	12.195	152	72407	42.551	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	85.100%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU071822\
Data File : VU049774.D
Acq On : 18 Jul 2022 20:46
Operator : SY/MD
Sample : N3700-16ME
Misc : 6.34g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
F5B57ME

Quant Time: Jul 19 02:14:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM071122WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Jul 19 02:09:49 2022
Response via : Initial Calibration

