

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU071824\
 Data File : VU060018.D
 Acq On : 18 Jul 2024 15:36
 Operator : MD/SY
 Sample : P3298-03
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BHAY7

Quant Time: Jul 19 01:34:20 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR071824WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Jul 19 01:30:22 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	144171	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	148904	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	65986	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	50013	4.986	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	99.800%
7) Chloroethane-d5	1.911	69	47270	5.073	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	101.400%
11) 1,1-Dichloroethene-d2	2.560	65	19576	4.998	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	100.000%
20) 2-Butanone-d5	4.624	46	127217	54.246	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	108.500%
24) Chloroform-d	5.055	84	102264	4.851	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.000%
26) 1,2-Dichloroethane-d4	5.695	65	51780	5.275	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	105.400%
32) Benzene-d6	5.721	84	203349	5.097	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.000%
36) 1,2-Dichloropropane-d6	6.685	67	66219	5.209	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	104.200%
41) Toluene-d8	7.891	98	167914	4.658	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.200%
43) trans-1,3-Dichloroprop...	8.177	79	21691	5.012	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	100.200%
46) 2-Hexanone-d5	8.627	63	102149	53.194	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	106.380%
56) 1,1,2,2-Tetrachloroeth...	10.749	84	59054	5.219	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	104.400%
66) 1,2-Dichlorobenzene-d4	12.187	152	66742	5.604	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	112.000%
Target Compounds						
13) Acetone	2.640	43	25273	16.291	ug/L	94
22) cis-1,2-Dichloroethene	4.660	96	15044	1.173	ug/L	97
34) Trichloroethene	6.544	95	6066	0.465	ug/L	95
47) Tetrachloroethene	8.547	164	79664	7.385	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU071824\
Data File : VU060018.D
Acq On : 18 Jul 2024 15:36
Operator : MD/SY
Sample : P3298-03
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BHAY7

Quant Time: Jul 19 01:34:20 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR071824WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Jul 19 01:30:22 2024
Response via : Initial Calibration

