

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU071824\
 Data File : VU060030.D
 Acq On : 18 Jul 2024 20:23
 Operator : MD/SY
 Sample : P3217-12
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 YDZH5

Quant Time: Jul 19 01:38:15 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR071824WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Jul 19 01:30:22 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	141466	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	145517	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	66396	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.592	65	43900	4.460	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	89.200%	
7) Chloroethane-d5	1.907	69	41977	4.591	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	91.800%	
11) 1,1-Dichloroethene-d2	2.554	65	19602	5.101	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	102.000%	
20) 2-Butanone-d5	4.653	46	126366	54.914	ug/L	0.02
Spiked Amount	50.000	Range 40 - 130	Recovery	=	109.820%	
24) Chloroform-d	5.055	84	92237	4.459	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	89.200%	
26) 1,2-Dichloroethane-d4	5.695	65	46683	4.847	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	97.000%	
32) Benzene-d6	5.717	84	189057	4.849	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	97.000%	
36) 1,2-Dichloropropane-d6	6.685	67	61380	4.941	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	98.800%	
41) Toluene-d8	7.891	98	155554	4.416	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	88.400%	
43) trans-1,3-Dichloroprop...	8.177	79	20079	4.748	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	95.000%	
46) 2-Hexanone-d5	8.634	63	98318	52.391	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	104.780%	
56) 1,1,2,2-Tetrachloroeth...	10.749	84	55108	4.984	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	99.600%	
66) 1,2-Dichlorobenzene-d4	12.187	152	62875	5.246	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	105.000%	
Target Compounds						Qvalue
17) Methyl tert-butyl Ether	3.364	73	67883	2.757	ug/L	99
22) cis-1,2-Dichloroethene	4.660	96	13174	1.047	ug/L	98
34) Trichloroethene	6.537	95	8595	0.674	ug/L	92
47) Tetrachloroethene	8.550	164	4998	0.474	ug/L	94
51) Chlorobenzene	9.444	112	15757	0.471	ug/L	96
67) 1,2-Dichlorobenzene	12.209	146	3753	0.168	ug/L #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU071824\
Data File : VU060030.D
Acq On : 18 Jul 2024 20:23
Operator : MD/SY
Sample : P3217-12
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
YDZH5

Quant Time: Jul 19 01:38:15 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR071824WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Jul 19 01:30:22 2024
Response via : Initial Calibration

