

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU071824\
 Data File : VU060042.D
 Acq On : 19 Jul 2024 01:33
 Operator : MD/SY
 Sample : VU0718WBL02
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK020

Quant Time: Jul 19 02:50:13 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR071824WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Jul 19 02:49:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	142221	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	145685	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	63832	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	50314	5.085	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	101.600%
7) Chloroethane-d5	1.911	69	47143	5.128	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	102.600%
11) 1,1-Dichloroethene-d2	2.563	65	19949	5.164	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	103.200%
20) 2-Butanone-d5	4.631	46	121577	52.552	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	105.100%
24) Chloroform-d	5.055	84	103202	4.962	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.200%
26) 1,2-Dichloroethane-d4	5.695	65	52124	5.383	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	107.600%
32) Benzene-d6	5.721	84	212359	5.441	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	108.800%
36) 1,2-Dichloropropane-d6	6.682	67	68402	5.500	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	110.000%
41) Toluene-d8	7.894	98	176942	5.017	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.400%
43) trans-1,3-Dichloroprop...	8.177	79	20782	4.909	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	98.200%
46) 2-Hexanone-d5	8.627	63	94663	50.385	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	100.780%
56) 1,1,2,2-Tetrachloroeth...	10.749	84	56124	5.070	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	101.400%
66) 1,2-Dichlorobenzene-d4	12.187	152	68629	5.956	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	119.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU071824\
Data File : VU060042.D
Acq On : 19 Jul 2024 01:33
Operator : MD/SY
Sample : VU0718WBL02
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK020

Quant Time: Jul 19 02:50:13 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR071824WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Jul 19 02:49:52 2024
Response via : Initial Calibration

