

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU071824\
 Data File : VU060045.D
 Acq On : 19 Jul 2024 08:44
 Operator : MD/SY
 Sample : P3217-19DL 100X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 YDZD6DL

Quant Time: Jul 20 00:54:57 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR071824WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Jul 19 02:49:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	153582	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	152002	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	76839	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	40838	3.822	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	76.400%
7) Chloroethane-d5	1.907	69	40500	4.080	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	81.600%
11) 1,1-Dichloroethene-d2	2.560	65	16590	3.976	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	79.600%
20) 2-Butanone-d5	4.624	46	109613	43.876	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	87.760%
24) Chloroform-d	5.055	84	89160	3.970	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	79.400%
26) 1,2-Dichloroethane-d4	5.695	65	43846	4.193	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	83.800%
32) Benzene-d6	5.721	84	180634	4.435	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.800%
36) 1,2-Dichloropropane-d6	6.682	67	58231	4.487	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	89.800%
41) Toluene-d8	7.891	98	151276	4.111	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	82.200%
43) trans-1,3-Dichloroprop...	8.177	79	20185	4.569	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	91.400%
46) 2-Hexanone-d5	8.627	63	86727	44.243	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	88.480%
56) 1,1,2,2-Tetrachloroeth...	10.749	84	49789	4.311	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	86.200%
66) 1,2-Dichlorobenzene-d4	12.187	152	64928	4.681	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	93.600%
Target Compounds						Qvalue
5) Vinyl chloride	1.599	62	37121	2.361	ug/L	100
22) cis-1,2-Dichloroethene	4.656	96	151552	11.096	ug/L	99
42) Toluene	7.962	91	162970	3.050	ug/L	99
52) Ethylbenzene	9.566	91	13226	0.248	ug/L	97
53) m,p-Xylene	9.688	106	18428	0.895	ug/L	96
54) o-Xylene	10.093	106	14511	0.737	ug/L	98
62) 1,3,5-Trimethylbenzene	11.084	105	14829	0.386	ug/L	96
63) 1,2,4-Trimethylbenzene	11.463	105	41710	1.124	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU071824\
Data File : VU060045.D
Acq On : 19 Jul 2024 08:44
Operator : MD/SY
Sample : P3217-19DL 100X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
YDZD6DL

Quant Time: Jul 20 00:54:57 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR071824WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Jul 19 02:49:52 2024
Response via : Initial Calibration

