




Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU071824\
Data File : VU060047.D
Acq On : 19 Jul 2024 09:29
Operator : MD/SY
Sample : P3237-08
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
E1GE6

Manual IntegrationsAPPROVED

Quant Time: Jul 20 00:55:30 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR071824WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Jul 19 02:49:52 2024
Response via : Initial Calibration

Reviewed By :Mahesh Dadoda 07/24/2024
Supervised By :Semsettin Yesilyurt 07/24/2024



TIC: VU060047.D\data.ms

(13) Acetone (T)

2.451min (-0.193) 459.24 ug/L

response 739468

Ion	Exp%	Act%
43.05	100.00	100.00
58.05	41.10	2.06
0.00	0.00	0.00
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU071824\
Data File : VU060047.D
Acq On : 19 Jul 2024 09:29
Operator : MD/SY
Sample : P3237-08
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
E1GE6

Manual IntegrationsAPPROVED

Reviewed By :Mahesh Dadoda 07/24/2024
Supervised By :Semsettin Yesilyurt 07/24/2024

Quant Time: Jul 20 00:55:30 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR071824WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Jul 19 02:49:52 2024
Response via : Initial Calibration



TIC: VU060047.D\data.ms

(13) Acetone (T)

2.644min (+ 0.000) 23.72 ug/L m

response 38202

Ion	Exp%	Act%
43.05	100.00	100.00
58.05	41.10	39.89
0.00	0.00	0.00
0.00	0.00	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU071824\
 Data File : VU060047.D
 Acq On : 19 Jul 2024 09:29
 Operator : MD/SY
 Sample : P3237-08
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 E1GE6

Manual IntegrationsAPPROVED

Quant Time: Jul 20 00:55:30 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR071824WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Jul 19 02:49:52 2024
 Response via : Initial Calibration

Reviewed By :Mahesh Dadoda 07/24/2024
 Supervised By :Semsettin Yesilyurt 07/24/2024

Compound	R.T.	Q	Ion	Response	Conc	Units	Dev(Min)

Internal Standards							
1) 1,4-Difluorobenzene	6.242	114		149643	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117		148749	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152		68848	5.000	ug/L	0.00
System Monitoring Compounds							
4) Vinyl Chloride-d3	1.596	65		38845	3.731	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	74.600%	
7) Chloroethane-d5	1.911	69		39405	4.074	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	81.400%	
11) 1,1-Dichloroethene-d2	2.563	65		16319	4.014	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	80.200%	
20) 2-Butanone-d5	4.624	46		108952	44.759	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	89.520%	
24) Chloroform-d	5.055	84		90699	4.145	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	82.800%	
26) 1,2-Dichloroethane-d4	5.695	65		44044	4.323	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	86.400%	
32) Benzene-d6	5.721	84		183588	4.607	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.200%	
36) 1,2-Dichloropropane-d6	6.682	67		57909	4.560	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	91.200%	
41) Toluene-d8	7.891	98		153927	4.275	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	85.400%	
43) trans-1,3-Dichloroprop...	8.174	79		17920	4.145	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	83.000%	
46) 2-Hexanone-d5	8.628	63		84024	43.801	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	87.600%	
56) 1,1,2,2-Tetrachloroeth...	10.750	84		50189	4.440	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	88.800%	
66) 1,2-Dichlorobenzene-d4	12.187	152		63176	5.084	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	101.600%	
Target Compounds							
13) Acetone	2.644	43		38202m	23.725	ug/L	
16) Methylene chloride	3.039	84		2280	0.163	ug/L	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU071824\
Data File : VU060047.D
Acq On : 19 Jul 2024 09:29
Operator : MD/SY
Sample : P3237-08
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
E1GE6

Manual IntegrationsAPPROVED

Reviewed By :Mahesh Dadoda 07/24/2024
Supervised By :Semsettin Yesilyurt 07/24/2024

Quant Time: Jul 20 00:55:30 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR071824WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Jul 19 02:49:52 2024
Response via : Initial Calibration

