

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU071824\
 Data File : VU060038.D
 Acq On : 18 Jul 2024 23:34
 Operator : MD/SY
 Sample : P3237-04
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 E1GB6

Quant Time: Jul 19 01:40:33 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR071824WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Jul 19 01:30:22 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	151438	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	154101	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	71189	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	42501	4.034	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	80.600%
7) Chloroethane-d5	1.911	69	41064	4.195	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	84.000%
11) 1,1-Dichloroethene-d2	2.560	65	17534	4.262	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	85.200%
20) 2-Butanone-d5	4.627	46	130388	52.931	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	105.860%
24) Chloroform-d	5.055	84	93434	4.219	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.400%
26) 1,2-Dichloroethane-d4	5.695	65	48257	4.680	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.600%
32) Benzene-d6	5.721	84	188096	4.556	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.200%
36) 1,2-Dichloropropane-d6	6.685	67	61445	4.670	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	93.400%
41) Toluene-d8	7.891	98	158246	4.242	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	84.800%
43) trans-1,3-Dichloroprop...	8.177	79	20156	4.501	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	90.000%
46) 2-Hexanone-d5	8.627	63	102653	51.654	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	103.300%
56) 1,1,2,2-Tetrachloroeth...	10.749	84	55330	4.725	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	94.600%
66) 1,2-Dichlorobenzene-d4	12.187	152	67225	5.232	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	104.600%
Target Compounds						Qvalue
13) Acetone	2.647	43	30491	18.712	ug/L	100
16) Methylene chloride	3.039	84	3306	0.233	ug/L	92


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU071824\
Data File : VU060038.D
Acq On : 18 Jul 2024 23:34
Operator : MD/SY
Sample : P3237-04
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
E1GB6

Quant Time: Jul 19 01:40:33 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR071824WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Jul 19 01:30:22 2024
Response via : Initial Calibration





#13

Acetone

Concen: 18.712 ug/L

RT: 2.647 min Scan# 41

Delta R.T. 0.003 min

Lab File: VU060038.D

Acq: 18 Jul 2024 23:34

Instrument :

MSVOA_U

ClientSampleId :





E1GB6

Tgt Ion: 43 Resp: 30491

Ion Ratio Lower Upper

43 100

58 41.0 0.0 82.2



#16

Methylene chloride

Concen: 0.233 ug/L

RT: 3.039 min Scan# 544

Delta R.T. 0.003 min

Lab File: VU060038.D

Acq: 18 Jul 2024 23:34

Tgt Ion: 84 Resp: 3306

Ion Ratio Lower Upper

84 100

86 56.1 44.4 82.5

49 94.9 72.0 133.6

