

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU071918\
 Data File : VU025515.D
 Acq On : 19 Jul 2018 12:10
 Operator : MD/SY
 Sample : J3984-17DL 5X
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 DB1A7DL

Quant Time: Jul 20 02:52:41 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM071718WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jul 20 02:44:53 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.89	114	321365	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	283542	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.49	152	124595	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	5458	3.60	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	7.20%#
7) Chloroethane-d5	1.69	69	2656	2.06	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	4.12%#
11) 1,1-Dichloroethene-d2	2.28	63	4406	1.32	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	2.64%#
21) 2-Butanone-d5	4.18	46	193	0.12	ug/L	0.01
Spiked Amount	100.000	Range	40 - 130	Recovery	=	0.12%#
24) Chloroform-d	4.65	84	708	0.18	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	0.36%#
26) 1,2-Dichloroethane-d4	5.31	65	643	0.24	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	0.48%#
32) Benzene-d6	5.35	84	2841	0.41	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	0.82%#
36) 1,2-Dichloropropane-d6	6.34	67	98	0.04	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	0.08%#
41) Toluene-d8	7.57	98	2565	0.39	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	0.78%#
43) trans-1,3-Dichloropropene-	7.86	79	269	0.23	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	0.46%#
47) 2-Hexanone-d5	8.32	63	301	0.26	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	0.26%#
57) 1,1,2,2-Tetrachloroethane-	10.44	84	676	0.20	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	0.40%#
64) 1,2-Dichlorobenzene-d4	11.86	152	1991	0.80	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	1.60%#

Target Compounds

					Ovalue
5) Vinyl chloride	1.41	62	26444	11.24	ug/L 99
17) trans-1,2-Dichloroethene	2.99	96	34320	16.66	ug/L 95
20) cis-1,2-Dichloroethene	4.23	96	129444	51.32	ug/L 97
34) Trichloroethene	6.19	95	45787	20.62	ug/L 96
46) Tetrachloroethene	8.23	164	3556	1.92	ug/L 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU071918\
Data File : VU025515.D
Acq On : 19 Jul 2018 12:10
Operator : MD/SY
Sample : J3984-17DL 5X
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
DB1A7DL

Quant Time: Jul 20 02:52:41 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM071718WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Jul 20 02:44:53 2018
Response via : Initial Calibration

