


Data Path : Z:\voasrv\HPCHEM1\MSVOA U\Data\VU071919\
 Data File : VU033331.D
 Acq On : 19 Jul 2019 11:08
 Operator : JC/SP
 Sample : BFB37
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BFB37

Integration File: LSCINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM071219WMA.M
 Title : VOC Analysis
 Last Update : Fri Jul 19 22:54:17 2019



AutoFind: Scans 2861, 2862, 2863; Background Corrected with Scan 2850

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.5	6429	PASS
75	95	30	80	50.5	17584	PASS
95	95	100	100	100.0	34832	PASS
96	95	5	9	6.7	2342	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	120	74.8	26037	PASS
175	174	5	9	8.5	2210	PASS
176	174	95	101	95.1	24770	PASS
177	176	5	9	6.7	1648	PASS

Scan 2850 (10.254 min): VU033331.D
BFB37

m/z	Abundance
39.90	175.0
44.00	431.0

Instrument :
MSVOA_U
ClientSampleId :
BFB37

m/z	Abundance
36.00	274.0
37.00	1867.0
38.00	1502.0
39.00	708.0
40.00	232.0
44.00	568.0
44.90	334.0
47.10	338.0
47.80	179.0
49.00	1309.0
50.00	6356.0
51.00	2013.0
51.90	150.0
56.00	520.0
57.00	924.0
59.90	302.0
61.00	1663.0
62.00	1713.0
63.00	1490.0
64.00	218.0
65.00	226.0
68.00	3432.0
69.00	3840.0
70.00	367.0
71.80	182.0
73.00	1775.0
74.00	6244.0
75.10	16608.0
76.00	1558.0
77.00	232.0
77.80	223.0
78.90	1345.0
79.90	414.0
80.90	1380.0
81.90	253.0
86.90	1340.0
87.90	1111.0
91.10	125.0
92.00	1018.0
93.00	1315.0
94.00	3995.0
95.00	33592.0
96.00	2316.0
96.90	108.0
103.80	104.0
105.90	167.0
115.80	149.0
116.80	180.0
117.90	161.0
118.90	144.0
128.80	113.0
129.80	127.0
140.90	292.0
142.90	465.0
144.80	122.0
154.90	101.0
171.90	260.0
173.90	24368.0
175.00	2043.0
176.00	23368.0
176.90	1550.0

Instrument :
MSVOA_U
ClientSampleId :
BFB37

m/z	Abundance
35.90	330.0
37.00	1873.0
38.00	1608.0
39.00	699.0
40.00	241.0
44.00	581.0
44.90	377.0
47.00	339.0
47.80	160.0
49.00	1388.0
50.00	6603.0
51.00	2204.0
51.90	133.0
56.00	594.0
57.00	1083.0
59.90	344.0
61.00	1657.0
62.00	1712.0
63.00	1600.0
64.00	191.0
65.00	269.0
68.00	3534.0
69.00	4068.0
70.00	381.0
71.90	198.0
73.00	1754.0
74.00	6630.0
75.10	18120.0
76.00	1771.0
77.00	261.0
77.80	230.0
78.90	1379.0
79.90	496.0
80.90	1446.0
81.90	271.0
86.90	1515.0
87.90	1296.0
90.80	121.0
92.00	1113.0
93.00	1543.0
94.00	4259.0
95.00	35888.0
96.00	2425.0
96.90	128.0
103.90	151.0
105.90	221.0
115.80	137.0
116.90	246.0
117.80	174.0
118.90	202.0
128.00	100.0
128.80	119.0
129.80	130.0
140.90	319.0
142.80	482.0
144.90	113.0
171.00	111.0
171.90	284.0
173.90	26912.0
175.00	2297.0
176.00	25512.0
176.90	1676.0

Instrument :
MSVOA_U
ClientSampleId :
BFB37

m/z	Abundance
35.90	369.0
37.00	1705.0
38.00	1576.0
39.00	677.0
40.00	251.0
44.00	529.0
45.00	431.0
47.00	360.0
47.70	149.0
49.00	1320.0
50.00	6328.0
51.00	2153.0
56.00	631.0
57.00	1138.0
59.90	359.0
61.00	1596.0
62.00	1517.0
63.00	1581.0
64.00	172.0
65.00	267.0
68.00	3336.0
69.00	3873.0
70.00	334.0
72.00	187.0
73.00	1581.0
74.00	6149.0
75.00	18024.0
76.00	1737.0
76.90	267.0
77.80	171.0
78.90	1277.0
79.90	515.0
80.90	1360.0
81.80	252.0
86.90	1469.0
87.90	1339.0
90.80	156.0
92.00	1013.0
93.00	1544.0
94.00	4028.0
95.00	35016.0
96.00	2286.0
97.00	119.0
103.90	208.0
105.80	234.0
115.80	110.0
116.80	301.0
117.80	142.0
118.90	219.0
128.00	104.0
129.90	110.0
140.90	327.0
142.80	474.0
171.00	297.0
174.00	26832.0
175.00	2290.0
176.00	25432.0
176.90	1720.0

Instrument :
MSVOA_U
ClientSampleId :
BFB37