

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU071921\
 Data File : VU044360.D
 Acq On : 19 Jul 2021 12:21
 Operator : SY/MD
 Sample : M3035-10
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PT-ADD-WS

Quant Time: Jul 20 01:21:05 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U071621DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Fri Jul 16 14:36:39 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Fluorobenzene	6.121	96	52266	1.000	ug/l	# 0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.639	95	19627	0.930	ug/l	0.00
Spiked Amount 1.000			Recovery	=	93.000%	
68) 1,2-Dichlorobenzene-d4	12.201	152	19496	0.948	ug/l	0.00
Spiked Amount 1.000			Recovery	=	95.000%	
Target Compounds						
						Qvalue
8) 1,1,2-Trichloro-1,2,2-...	2.588	101	505467	36.944	ug/l	98
24) tert-Butyl Alcohol	3.202	59	31460	47.338	ug/l #	1
25) Methyl tert-Butyl Ether	3.378	73	1451361	38.954	ug/l	93
31) Isopropyl Ether	4.012	45	1144038	24.802	ug/l	94
32) Ethyl-t-butyl ether	4.523	59	1936432	42.574	ug/l	98
33) Tert-Amyl methyl ether	5.960	73	1581227	39.094	ug/l	98
73) n-propylbenzene	10.912	120	526146	26.939	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU071921\
Data File : VU044360.D
Acq On : 19 Jul 2021 12:21
Operator : SY/MD
Sample : M3035-10
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PT-ADD-WS

Quant Time: Jul 20 01:21:05 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U071621DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Fri Jul 16 14:36:39 2021
Response via : Initial Calibration

