

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU071921\
 Data File : VU044362.D
 Acq On : 19 Jul 2021 15:01
 Operator : SY/MD
 Sample : M3035-09DL 5X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PT-THM-WSDL

Quant Time: Jul 20 01:21:23 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U071621DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Fri Jul 16 14:36:39 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Fluorobenzene	6.121	96	51013	1.000	ug/l	# 0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.639	95	19569	0.950	ug/l	0.00
Spiked Amount 1.000			Recovery	=	95.000%	
68) 1,2-Dichlorobenzene-d4	12.201	152	18712	0.932	ug/l	0.00
Spiked Amount 1.000			Recovery	=	93.000%	
Target Compounds						
						Qvalue
15) Methylene Chloride	3.054	84	51697	2.796	ug/l	93
27) Chloroform	5.099	83	225813	8.621	ug/l	99
44) Bromodichloromethane	7.115	83	140628	7.486	ug/l	99
55) Dibromochloromethane	8.819	129	91205	7.220	ug/l	100
67) Bromoform	10.298	173	30267	4.281	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU071921\
Data File : VU044362.D
Acq On : 19 Jul 2021 15:01
Operator : SY/MD
Sample : M3035-09DL 5X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PT-THM-WSDL

Quant Time: Jul 20 01:21:23 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U071621DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Fri Jul 16 14:36:39 2021
Response via : Initial Calibration

