

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU071921\
 Data File : VU044363.D
 Acq On : 19 Jul 2021 15:35
 Operator : SY/MD
 Sample : M3035-10DL 5X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PT-ADD-WSDL

Manual Integrations
 APPROVED

Reviewed By : John Carlone 07/23/2021
 Supervised By : Mahesh Dadoda 07/23/2021

Quant Time: Jul 20 01:21:35 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U071621DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Fri Jul 16 14:36:39 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Fluorobenzene	6.121	96	53801	1.000	ug/l	# 0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.639	95	20271	0.934	ug/l	0.00
Spiked Amount 1.000			Recovery	=	93.000%	
68) 1,2-Dichlorobenzene-d4	12.201	152	19753	0.933	ug/l	0.00
Spiked Amount 1.000			Recovery	=	93.000%	
Target Compounds						
						Qvalue
8) 1,1,2-Trichloro-1,2,2-...	2.591	101	110646	7.856	ug/l	98
15) Methylene Chloride	3.054	84	61292	3.143	ug/l	93
24) tert-Butyl Alcohol	3.208	59	5129m	7.498	ug/l	
25) Methyl tert-Butyl Ether	3.378	73	321612	8.386	ug/l	92
31) Isopropyl Ether	4.012	45	248804	5.240	ug/l	95
32) Ethyl-t-butyl ether	4.523	59	428086	9.143	ug/l	98
33) Tert-Amyl methyl ether	5.960	73	320400	7.696	ug/l	99
73) n-propylbenzene	10.912	120	112554	5.599	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU071921\
Data File : VU044363.D
Acq On : 19 Jul 2021 15:35
Operator : SY/MD
Sample : M3035-10DL 5X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PT-ADD-WSDL

Manual Integrations
APPROVED

Reviewed By : John Carlone 07/23/2021
Supervised By : Mahesh Dadoda 07/23/2021

Quant Time: Jul 20 01:21:35 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U071621DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Fri Jul 16 14:36:39 2021
Response via : Initial Calibration

