

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU072018\
 Data File : VU025563.D
 Acq On : 20 Jul 2018 23:32
 Operator : MD/SY
 Sample : J4097-16
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 H0AB3

Quant Time: Jul 21 03:40:05 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM072018WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Jul 21 02:53:20 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.89	114	315376	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	285595	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.49	152	130340	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	77366	42.97	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	85.94%
7) Chloroethane-d5	1.68	69	66622	44.86	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	89.72%
11) 1,1-Dichloroethene-d2	2.27	63	128972	34.41	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	68.82%
21) 2-Butanone-d5	4.17	46	138621	99.73	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	99.73%
24) Chloroform-d	4.65	84	181572	45.27	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.54%
26) 1,2-Dichloroethane-d4	5.31	65	126237	47.60	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.20%
32) Benzene-d6	5.35	84	354999	48.47	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.94%
36) 1,2-Dichloropropane-d6	6.34	67	115519	48.95	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	97.90%
41) Toluene-d8	7.57	98	326430	46.47	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	92.94%
43) trans-1,3-Dichloropropene-	7.85	79	53967	44.77	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	89.54%
47) 2-Hexanone-d5	8.31	63	102177	103.39	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	103.39%
57) 1,1,2,2-Tetrachloroethane-	10.44	84	156522	47.42	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	94.84%
64) 1,2-Dichlorobenzene-d4	11.86	152	130823	49.52	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	99.04%

Target Compounds

					Ovalue
13) Acetone	2.32	43	6379	3.91 ug/L	94
18) Methyl tert-butyl Ether	3.00	73	19061	2.86 ug/L #	89
19) 1,1-Dichloroethane	3.45	63	35363	9.07 ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU072018\
Data File : VU025563.D
Acq On : 20 Jul 2018 23:32
Operator : MD/SY
Sample : J4097-16
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
H0AB3

Quant Time: Jul 21 03:40:05 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM072018WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Jul 21 02:53:20 2018
Response via : Initial Calibration

