

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU072022\
 Data File : VU049798.D
 Acq On : 20 Jul 2022 11:02
 Operator : SY/MD
 Sample : VU0720WBL01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK163

Quant Time: Jul 21 04:22:41 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM071122WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Jul 21 04:19:56 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	170366	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	166114	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	70291	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	71606	49.814	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	99.620%
7) Chloroethane-d5	1.906	69	54584	52.533	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	105.060%
11) 1,1-Dichloroethene-d2	2.562	63	91959	36.217	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	72.440%
21) 2-Butanone-d5	4.623	46	139233	113.464	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	113.460%
24) Chloroform-d	5.063	84	133542	52.452	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	104.900%
26) 1,2-Dichloroethane-d4	5.703	65	93883	54.485	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	108.960%
32) Benzene-d6	5.726	84	250236	49.586	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.180%
36) 1,2-Dichloropropane-d6	6.690	67	90222	53.152	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	106.300%
41) Toluene-d8	7.899	98	207580	47.234	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	94.460%
43) trans-1,3-Dichloroprop...	8.179	79	36181	50.591	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	101.180%
47) 2-Hexanone-d5	8.632	63	68122	89.218	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	89.220%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	124827	52.557	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	105.120%
66) 1,2-Dichlorobenzene-d4	12.192	152	70884	53.485	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	106.980%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU072022\
Data File : VU049798.D
Acq On : 20 Jul 2022 11:02
Operator : SY/MD
Sample : VU0720WBL01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK163

Quant Time: Jul 21 04:22:41 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM071122WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Jul 21 04:19:56 2022
Response via : Initial Calibration

