

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU072023\
 Data File : VU054826.D
 Acq On : 20 Jul 2023 11:13
 Operator : MD/SY
 Sample : VSTD0.0542
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VSTD0.05042

Quant Time: Jul 21 00:37:08 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM072023.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Jul 21 00:36:35 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.258	114	3815	0.500	ug/L	0.00
5) Chlorobenzene-d5	9.419	117	5639	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.808	152	1751	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.597	65	256	0.067	ug/L	0.00
4) 1,2-Dichloroethane-d4	5.709	65	122	0.060	ug/L	0.02
7) 1,2-Dichloropropane-d6	6.687	67	139	0.055	ug/L	0.00
8) Toluene-d8	7.887	98	459	0.061	ug/L	0.00
10) 1,1,2,2-Tetrachloroeth...	10.745	84	113	0.058	ug/L	0.00
Target Compounds						
						Qvalue
3) Vinyl chloride	1.597	62	386	0.066	ug/L #	72
6) Trichloroethene	6.544	95	155	0.056	ug/L	90
9) 1,2-Dibromoethane	8.924	107	85	0.050	ug/L #	83
12) 1,2,3-Trichloropropane	10.814	75	80	0.054	ug/L	91
13) 1,2-Dibromo-3-chloropr...	12.990	75	14	0.050	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU072023\
Data File : VU054826.D
Acq On : 20 Jul 2023 11:13
Operator : MD/SY
Sample : VSTD0.0542
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VSTD0.05042

Quant Time: Jul 21 00:37:08 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM072023.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jul 21 00:36:35 2023
Response via : Initial Calibration

