

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU072023\
 Data File : VU054827.D
 Acq On : 20 Jul 2023 11:37
 Operator : MD/SY
 Sample : VSTD0.143
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VSTD0.1043

Quant Time: Jul 21 00:37:18 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM072023.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Jul 21 00:36:35 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.258	114	3589	0.500	ug/L	0.00
5) Chlorobenzene-d5	9.418	117	5428	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.808	152	1676	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.597	65	396	0.110	ug/L	0.00
4) 1,2-Dichloroethane-d4	5.709	65	210	0.109	ug/L	0.02
7) 1,2-Dichloropropane-d6	6.687	67	248	0.102	ug/L	0.00
8) Toluene-d8	7.887	98	753	0.104	ug/L	0.00
10) 1,1,2,2-Tetrachloroeth...	10.745	84	196	0.104	ug/L	0.00
Target Compounds						
						Qvalue
3) Vinyl chloride	1.597	62	596	0.108	ug/L #	82
6) Trichloroethene	6.544	95	262	0.099	ug/L	92
9) 1,2-Dibromoethane	8.924	107	169	0.103	ug/L #	87
12) 1,2,3-Trichloropropane	10.814	75	126	0.089	ug/L #	88
13) 1,2-Dibromo-3-chloropr...	12.990	75	26	0.096	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU072023\
Data File : VU054827.D
Acq On : 20 Jul 2023 11:37
Operator : MD/SY
Sample : VSTD0.143
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VSTD0.1043

Quant Time: Jul 21 00:37:18 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM072023.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jul 21 00:36:35 2023
Response via : Initial Calibration

