

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU072023\
 Data File : VU054832.D
 Acq On : 20 Jul 2023 14:22
 Operator : MD/SY
 Sample : VU0720WBL01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK018

Quant Time: Jul 21 01:18:06 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM072023.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Jul 21 01:10:23 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.258	114	3678	0.500	ug/L	0.00
5) Chlorobenzene-d5	9.419	117	5441	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.808	152	1648	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.597	65	1623	0.440	ug/L	0.00
Spiked Amount 0.500	Range 40 - 130		Recovery =	88.000%		
4) 1,2-Dichloroethane-d4	5.685	65	929	0.470	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery =	94.000%		
7) 1,2-Dichloropropane-d6	6.687	67	1220	0.499	ug/L	0.00
Spiked Amount 0.500	Range 60 - 140		Recovery =	100.000%		
8) Toluene-d8	7.888	98	3347	0.462	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery =	92.000%		
10) 1,1,2,2-Tetrachloroeth...	10.746	84	977	0.517	ug/L	0.00
Spiked Amount 0.500	Range 65 - 120		Recovery =	104.000%		
Target Compounds						
3) Vinyl chloride	1.597	62	136	0.024	ug/L	Qvalue # 1

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU072023\
Data File : VU054832.D
Acq On : 20 Jul 2023 14:22
Operator : MD/SY
Sample : VU0720WBL01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK018

Quant Time: Jul 21 01:18:06 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM072023.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jul 21 01:10:23 2023
Response via : Initial Calibration

