

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU072122\
 Data File : VU049819.D
 Acq On : 21 Jul 2022 10:17
 Operator : SY/MD
 Sample : VU0721MBL01
 Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK164

Quant Time: Jul 22 04:11:06 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM071122WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jul 22 03:08:56 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	180182	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	168790	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	70214	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	66379	43.662	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	87.320%	
7) Chloroethane-d5	1.903	69	52956	48.189	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	96.380%	
11) 1,1-Dichloroethene-d2	2.562	63	83135	30.958	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	61.920%	
21) 2-Butanone-d5	4.620	46	148320	114.285	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	114.280%	
24) Chloroform-d	5.060	84	134277	49.868	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	99.740%	
26) 1,2-Dichloroethane-d4	5.700	65	95916	52.632	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	105.260%	
32) Benzene-d6	5.726	84	246810	48.132	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	96.260%	
36) 1,2-Dichloropropane-d6	6.690	67	89611	51.955	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	103.920%	
41) Toluene-d8	7.896	98	199211	44.611	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	89.220%	
43) trans-1,3-Dichloroprop...	8.179	79	36701	50.505	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	101.000%	
47) 2-Hexanone-d5	8.632	63	72686	93.686	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	93.690%	
56) 1,1,2,2-Tetrachloroeth...	10.755	84	128332	53.176	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	106.360%	
66) 1,2-Dichlorobenzene-d4	12.192	152	69226	52.292	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	104.580%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU072122\
Data File : VU049819.D
Acq On : 21 Jul 2022 10:17
Operator : SY/MD
Sample : VU0721MBL01
Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK164

Quant Time: Jul 22 04:11:06 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM071122WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Jul 22 03:08:56 2022
Response via : Initial Calibration

